

# Rigel UNI-SiM

# The world's smallest integrated NiBP, SpO2 and patient simulator.

The UNI-SiM is a handheld and battery-operated vital signs simulator capable of undertaking six synchronized vital signs parameters. This enables medical device engineers to quickly, easily and accurately perform NiBP, SpO2, ECG, temperature, IBP and respiration functionality tests simultaneously, using a single portable instrument.

With a fast boot up and single button simulation to repeat the last simulation value in seconds, the UNI-SiM reduces the time taken to test the correct performance of a wide range of medical devices and equipment.

It is easy-to-use and incorporates the full functionality of conventional NiBP and SpO2 simulators with a comprehensive patient simulator.

Compatibility with the Rigel PULS-R universal SpO2 simulation finger creates a truly versatile and valuable tool for every biomed engineer in need of a cost-effective and lightweight solution for testing the performance of vital signs monitors.

# **Key Features**

- Compact and cost-effective
- 6-in-1 vital signs simulation
- Fast start up and single button simulation
- Accurate and real-life simulations
- User definable NiBP simulations
- On-board automation and data storage
- User programmable patient conditions
- Universal SpO2 simulation with PULS-R
- Easy and accurate probe placement with PULS-R

#### **Simulation Functions**

- NiBP (systolic and diastolic)
- ECG
- Respiration
- SpO2
- IBP
- Temperature

## **End User Types**

BMETs requiring a cost-effective solution for carrying out performance checks and full PMs on vital signs monitors.









# Compact and cost-effective

A highly cost-effective and compact solution for testing 6 of the most common vital signs using a single battery-powered simulator.





## 6-in-1 vital signs simulation

The UNI-SiM incorporates full NiBP simulation including dynamic and static pressure simulation and leak and over-pressure testing with a comprehensive SpO2 and ECG simulator, all in a single hand-held enclosure.

## Fast start up and single button simulation

Automatic power-up of the most recent settings provides simulation of all 6 vital signs with the press of a single button, saving valuable time when setting up the simulator.





.com

深圳为尔康科技有限公司 联系人:曾祥满 手机:13632925349 QQ: 274798107 电话: 0755-28896837 地址: 深圳市龙岗区沙平北路111号608A 网址:www.medicalQC.com 邮箱:szchina1718@163.com







## Accurate and real-life simulations

Fully synchronized simulation signals provide the closest and most accurate representation of a real patient.

## User definable NiBP simulations

User configurable and physiologically correct systolic and diastolic pressures provide a truly universal and accurate NiBP simulator.





## On-board automation and data storage

Simple record management for up to 5,000 asset records. The simulated values from the patient monitor can be entered directly into the UNI-SiM using the built-in keyboard, improving traceability and reducing the need for manual data recording.





深圳为尔康科技有限公司 联系人:曾祥满 手机:13632925349
QQ:274798107 电话:0755-28896837 地址:深圳市龙岗区沙平北路111号608A
网址:www.medicalQC.com 邮箱:szchina1718@163.com



## User programmable patient conditions

Patient specific physiological conditions can be created and stored in the UNI-SiM to provide a highly customisable simulator. A truly versatile tool capable of meeting even the most demanding test protocols.





# Universal SpO2 simulation with PULS-R

Reduce the need for separate accessories with the universal PULS-R SpO2 simulation finger.

This compact SpO2 simulation enables accurate SpO2 simulations in 1% resolution from as low as 30%\* using the pre-programmed manufacturer specific R-curves.\*subject to monitor capability

## **Easy and accurate probe placement with PULS-R** Unique probe placement LED's ensure accurate and correct simulation for each type of SpO2 probe.

The Rigel PULS-R has status LEDs which light up to indicate whether a probe connection has been achieved.



深圳为尔康科技有限公司 联系人:曾祥满 手机:13632925349
QQ:274798107 电话:0755-28896837 地址:深圳市龙岗区沙平北路111号608A
网址:www.medicalQC.com 邮箱:szchina1718@163.com





# **Technical Specifications - UNI-SiM**

### Non-Invasive Blood Pressure Simulation

Waveform Pulse Volume Heart Rate Integrated Pump Leak Test Chronometer Digital Manometer Pressure Accuracy Pressure Units

Oscillometric High, Medium, Low, Paediatric 20 - 300BPM 0 to 350mmHg user configurable User configurable between 0-350mmHg Configurable up to 999 secs 0 - 410mmHg +/- 0.5% FS mmHg, inHg, kg/cm2, cmH2O, mBar, PSI, in H2O and kPa

#### Oxygen Saturation Simulation (SpO2 Adapter boxes)

Simulation	Optronic (Electronic & Optical)	
Simulation via	Probe & full chain	
Range	50 to 100%	
Accuracy	± 0.5% of reading between 80-100% SpO2	
	± 1% of reading between 50-79% SpO2	
Heart Rate	20-300BPM	
Accuracy	± 1BPM	
Compatibility	GE - Datex, Nellcor, N.Oximax, Masimo,	
	Datascope, Mindray, Nonin, Philips (HP)	
Perfusion Index	-20% (5% default)	
Chronometer	Test option to test response time	
	SpO2 monitor.	

## Oxygen Saturation Simulation (PULS-R)

Range	30 t	o 100%			
Repeatability	±	5%**	of	reading	between
	30-	59% Sp(	D2		
	± 3% of reading between 60-99% SpO2				
	±	3%	of	reading	between
90-100% SpO2					
Accuracy of simulation when used wit	th the c	orrespondir	na R-cur	ves	

\*Based on using the same probe and monitor setup

\*\*Note that some monitor types might not be able to display low range sats

Heart Rate 30-300BPM\*\*\* ± 1BPM Accuracy Compatibility Beijing Choice, Criticare, GE Tuffsat, Masimo, Mindray, Nellcor, Nellcor Oximax, Nihon Kohden, Nonin, Novametrix, Philips / HP

\*\*\*Subject to monitor capability

#### ECG Arrhythmia Simulator

Simulation

Heart Rate

Accuracy Amplitudes

Accuracy

5 lead simulation including high level output on Normal Sinus Rhythm (NSR), ST Elevation. ST Depression. Myocardial Infarcation, Tall T 20 - 300BPM ±1BPM 0.5/1/1.5/2/2.5/3/3.5/4/4.5/5mV ± 2% Connection High-Level ECG 3.5mm jack plug

#### ST Elevation / Depression

Heart Rate Elevation % Elevation Slope 20 - 300BPM 7%, 13%, 20% Positive, Negative, Flat

#### Myocardial Infarction

Туре	Ischemia, Injury, Infarction, Inferior Infarction
Heart Rate	20 – 300BPM
Tall T	
Heart Rate	80BPM

Heart Rate T Wave Amplitude

0 - 1.2mV (steps of 0.1mV)

Arrhythmia Waveforms Full 12 lead simulation Simulation Amplitudes 0.5 / 1 / 1.5 / 2 / 2.5 / 3 / 3.5 / 4 / 4.5 / 5mV

Heart Rate (where applicable) 20 - 300BPM

#### Atrial

Sinus Arrhythmia (SA), Missing Beat, Atrial Flutter (AFLT), Atrial Fibrillation (AFB), Paroxysmal Atrial Tachycardia (PAT), Junctional Premature Contraction

#### Atrial Conduction

First Degree AV Block, Second Degree AV Block - Mobitz I, Second Degree AV Block - Mobitz II, Third Degree AV Block, Right Bundle Branch Block (RBB), Left Bundle Branch Block (LBB), Left Anterior Hemiblock

#### Ventricular

Shape

Premature Ventricular Contraction - Intermittent Premature Ventricular Contraction - Continuous, Bigeminy, Trigeminy, Ventricular Flutter (VFLT), Ventricular Fibrillation Fine (VFBF), Ventricular Fibrillation Coarse (VFBC), Monomorphic Ventricular Tachycardia (MVT), Polymorphic Ventricular Tachycardia (PVT), Right Focal (PVC)

Sine, Square, Triangle, and Pulse

3.5 /

Atrial,

#### Performance Waveforms

Rates	0.1 to 0.9Hz (in steps of 0.1) 1 to 100Hz (in steps of 1) 0.5 / 1 / 1.5 / 2 / 2.5 / 3 / 3.5 /
Pulse	4 / 4.5 / 5mV 1mV, 4 sec delay (20ms pulse duration)
Pacer Waveforms	
Available	Synchronous Atrial, Asynchronous Atrial, Paver Only, Ventricular Pacer, Atrial & Ventricular Pacer
QRS	1mV
Pacer Pulse Amplitude	0.1 – 2mV
Pacer Pulse Polarity	Positive, Negative
Pacer Pulse Width	0.1 – 2ms

70BPM

#### **R** Wave Detection

Heart Rate R Wave Width

## **Temperature Simulation**

Simulation Range Accuracy Default Setting YSI 400 / 700 Static Preset at 25, 33, 37 and 41°C ± 0.1 °C YSI 400 37°C

10 - 120ms (steps of 10ms)





深圳为尔康科技有限公司 联系人:曾祥满 手机:13632925349

QQ: 274798107 电话: 0755-28896837 地址: 深圳市龙岗区沙平北路111号608A

网址:www.medicalQC.com 邮箱:szchina1718@163.com



# **Technical Specifications - UNI-SiM (Continued)**

## **Respiration Simulation**

Rates

Base Resistances Accuracy Resistance Variations Accuracy Default Settings Apnoea Simulation 5, 10, 15, 30, 60, 120, 180 Breaths per Minute 250, 500, 750, 1000 $\Omega$  $\pm$  5% 0.1, 0.5, 1.0, 1.5 $\Omega$  $\pm$  10% 15BPM / 250 $\Omega$  / 0.1 $\Omega$ 0 – 60 seconds duration 0 – 300 seconds interval

## **General Specifications - UNI-SiM**

Operation Battery Charger Supply Battery Life

Memory Capacity

Communication

Size (L x W x D)

Operating Conditions

Storage Environment Environmental Protection

Display

Keypad

Weight

Battery cell, in-situ charge 100-240VAC, 50/60Hz 12VDC centre positive 8 hours standby or a maximum of 200 NiBP simulations Approx. 5,000 records via Bluetooth Monochrome, ¼ VGA full graphics Alpha-numeric <1.5kg, <3.5lbs 270 x 110 x 75mm / 10.5 x 4 x 3" 10-30°C, 0-90% RH - NC -15° - +60°C IP 40

## Invasive Blood Pressure Simulation

Channels Static Dynamic Accuracy Excitation Voltage Impedance Simulated Sensitivity 2 channels 0 to 300mmHg 0-300mmHg for Systolic & Diastolic  $\pm$  1mmHg 2 - 16V 350 $\Omega$  Nominal 5 $\mu$ V / V / mmHg

#### Service & Warranty

UNI-SiM comes with a free upgraded 24 month warranty (subject to terms and conditions, available at www.rigelmedical.com/register-product)

#### Standard Accessories (supplied with UNI-SiM)

Standard Accessories (supplied with PULS-R)

PULS-R universal SpO2 simulation finger

Carry case

- NiBP tubing kit
- ECG adaptor module

## **Optional Accessories**

- IBP connect cables
- NiBP accessories
- Temperature connect cablesECG cables and leads

ECG snap-on adaptors

Quick start guide

Power supply

To find out more, visit www.rigelmedical.com/sim-accessories

## **Specifications - PULS-R**

#### Supported Default R Curves

Beijing Choice	
GE Tuffsat	
Mindray	
Nellcor Oximax	
Nonin	
Philips / HP	

Heart Rate Setting

30-300BPM (subject to Monitor Compatibility)

Accuracy of Simulation when used with the Corresponding R Curves				
Resolution	Range	Repeatability*		
1% steps	30-59%	±5%**		
1% steps	60-89%	±3%		
1% steps	90-100%	±1%		

Criticare

Masimo Nellcor Nihon Kohden Novametrix

Part Numbers UNI-SiM PULS-R

370A930 399A910

Rev 2, 2015



深圳为尔康科技有限公司 联系人:曾祥满 手机:13632925349

QQ: 274798107 电话: 0755-28896837 地址: 深圳市龙岗区沙平北路111号608A

网址:www.medicalQC.com 邮箱:szchina1718@163.com

 Quick Start Guide